

Accuracy and repeatability of quantitative fluoroscopy for the measurement of sagittal plane translation and finite centre of rotation in the lumbar spine

Alexander Breen ^a, Alan Breen ^b,

^a Institute for Musculoskeletal Research and Clinical Implementation, Anglo-European College of Chiropractic, 13-15 Parkwood Road, Bournemouth, Dorset BH5 2DF, UK

^b School of Design Engineering and Computing, Bournemouth University, Talbot Campus, Poole, Dorset BH12 5BB, UK

Abstract

Quantitative fluoroscopy (QF) was developed to measure intervertebral mechanics in vivo and has been found to have high repeatability and accuracy for the measurement of intervertebral rotations. However, sagittal plane translation and finite centre of rotation (FCR) are potential measures of stability but have not yet been fully validated for current QF. This study investigated the repeatability and accuracy of QF for measuring these variables. Repeatability was assessed from L2-S1 in 20 human volunteers. Accuracy was investigated using 10 consecutive measurements from each of two pairs of linked and instrumented dry human vertebrae as reference; one which tilted without translation and one which translated with- out tilt. The results found intra- and inter-observer repeatability for translation to be 1.1 mm or less (SEM) with fair to substantial reliability (ICC 0.533–0.998). Intra-observer repeatability of FCR location for inter-vertebral rotations of 5 ° and above ranged from 1.5 mm to 1.8 mm (SEM) with moderate to substantial reliability (ICC 0.626–0.988). Inter-observer repeatability for FCR ranged from 1.2 mm to 5.7 mm, also with moderate to substantial reliability (ICC 0.621–0.878). Reliability was substantial (ICC > 0.81) for 10/16 measures for translation and 5/8 for FCR location. Accuracy for translation was 0.1 mm (fixed centre) and 2.2 mm (moveable centre), with an FCR error of 0.3 mm(x) and 0.4 mm(y) (fixed centre). This technology was found to have a high level of accuracy and with a few exceptions, moderate to substantial repeatability for the measurement of translation and FCR from fluoroscopic motion sequences.

Introduction

The in vivo measurement of intervertebral motion in the lumbar spine in individuals has been progressing. This information has traditionally been obtained as displacement on flexion-extension radiographs, however, this has been consistently found to be prone to large errors and variability between observers [1–5] . The method also suffers from the inability to detect the true end-range during motion and lack of standardised measurement methods [6] . Studies of quantitative fluoroscopy (QF) for measuring lumbar spine intervertebral kinematics using continuous motion tracking began in the 1980s [7] . QF measures continuous intervertebral motion and extracts end of range measurement from wherever it occurs in the bending sequence, giving a radiation dose similar to a conventional radiographic examination [8,9] . Various iterations have been found to have good

repeatability and accuracy for measuring intervertebral rotations at lumbar and cervical levels [5,9–12]. However, excessive translation is thought to be more closely associated with back symptoms [13]. Translation also affects the finite centre of rotation (FCR) and the latter is an expression of the distribution of loading between the disc and facets during upright flexion-extension motion [14]. It is also said that the centre of reaction force (CR) can be extrapolated from the FCR [14]. QF technology employs standardised image registration and analysis protocols with relatively straightforward and inexpensive hardware in contrast to specialist MR, CT or dual fluoroscopic systems which are not as readily available in hospital settings. However, the literature addressing the repeatability and accuracy of translation and FCR measurement from fluoroscopy is based on different techniques. For example, Cerciello et al. determined the accuracy of measuring intervertebral rotation and FCR location in 2-D using stepped positions in a calibration specimen rather than from continuous motion [15]. Wang et al. and Lin et al. determined the accuracy of translation measurement in ovine specimens using 2D-3D dual fluoroscopic systems where the geometry was informed by magnetic resonance or CT-based vertebral models of the same participant rather than a calibrated reference [16,17]. These studies also found excellent accuracy—and in the case of Wang et al. good repeatability—for translation measurement. However, they involved greater radiation dose and expense, while Yeager et al. found good repeatability for pooled vertebral levels using a less elaborate low-dose 2-D clinical QF system, but did not assess levels individually [5,18]. The validation of QF technology for in vivo translation and FCR measurement from continuous motion sequences is therefore incomplete. The aim of this study was to determine the current accuracy and repeatability of 2-D QF for measuring lumbar intervertebral translation and FCR location during motion using a standardised patient motion protocol. This research involved the use of two calibrated human cadaveric specimens to assess accuracy during sagittal plane motion in a prescribed pathway and repeatability in 20 volunteers executing a standardised bending protocol.




Fig. 1. Lumbar intervertebral motion specimens. (A) Fixed centre specimen. (B) Movable centre specimen.

2. Methods

2.1. Accuracy study

Two sets of dry cadaveric vertebral pairs were used to provide reference data. Specimen A (Fig. 1 A) consisted of L4 and L5 vertebrae joined at their end-plate centres by a universal joint 4 mm high, representing a fixed centre of rotation with zero translation. Specimen B (Fig. 1 B) comprised of L3

and L4 vertebrae. These were joined at their end-plate centres by a plastic linkage which allowed translation of the upper vertebra without rotation. It was driven by an actuator motor and controller (Arduino Software Ltd., UK—resolution 0.01 mm) providing anterior to posterior translation across the lower vertebral end-plate during the rotation. Both specimens were mounted on rigid bases and positioned 15 cm from a motion frame which incorporated a rotating disc (Fig. 1 A and B). The central ray of a C-arm digital fluoroscope (Siemens Arcadis Avantic—Siemens GMBH, Germany) was positioned so as to pass through the centre of the disc space. A block of animal soft tissue was interposed between the X-ray source, the models and the fluoroscope's image intensifier to degrade the images by generating soft tissue scatter. The superior vertebra of specimen A was rotated to 18° of flexion and return representing an arbitrary physiological maximum measured using a tilt sensor (Axminster instruments UK—resolution $\pm 0.002^\circ$) [19]. This was done using a rod driven by a vertical rotating disc embedded in a vertical motion frame (Fig. 1 A). It was controlled and driven by a laptop computer using bespoke software (Daqfactory VSC). The superior vertebra of Specimen B was translated posteriorly across 50% of the lower vertebral end-plate and back again. This was an arbitrary range designed to allow direct comparison between the reference and index values, which should apply, within reason, no matter how large or small the translation. Rotation was at 3 °/s and translation at 1.5 mm/s. These procedures were repeated 10 times for each specimen. Images were recorded at 15 frames per second during the 10 sequences for each specimen. All image sequences were analysed by one trained observer.

2.2. Repeatability study

Data were obtained from a parallel study of twenty volunteers being examined for passive recumbent lumbar motion [9]. These were recruited using the eligibility criteria described in Table 1 and following a favourable opinion from the National Research Ethics Service (REC reference 0/H0502/99). Each participant was positioned in the lateral decubitus position on a horizontal motion frame with the central ray of the fluoroscope positioned to pass through the L4 vertebra (Fig. 2). The inferior section of the motion frame was rotated through 40 ° of flexion over a 12 s interval using the motion controller (Daqfactory VSC—Heatherose Electronics Ltd., UK). This was immediately followed by 40 ° of extension. The effective radiation dose for this procedure has been estimated as 0.24 mSv [18]. After transfer of images from the fluoroscope to an image processing workstation, two trained observers (a senior radiographer and a medical physicist) analysed the same 40 image sequences for inter-observer repeatability (two sequences per participant for the 20 participants). Five repeated mark-ups of flexion and extension images of intervertebral levels from L2-S1 took approximately 20 min. Observers were blinded to each other's image registrations. The second observer also analysed each image sequence twice for intra-observer repeatability.

2.3. Kinematic data extraction

The fluoroscopic sequences were transferred to a desktop computer and Image J (v 1.47 for Windows OS) was used to separate the individual images from the digital sequences. The images underwent user defined edge enhancement, after which templates were manually placed five times around each vertebral body (L2-S1) in the first image. Bespoke software written in Matlab (V R2007b, The Mathworks Inc.) used a cross-correlation method to obtain automated frame to frame image tracking of the vertebral bodies in subsequent images [20]. Co-ordinates were placed on the vertebral body corners in the first image, linked to the tracking templates and used to register the

vertebrae in two dimensional space in each frame. Tracking was verified for quality assurance by viewing all sequences and repeating any tracking that failed. The displacements between each pair of tracked positions were calculated using Distortion Compensated Radiographic Analysis [21]. These were averaged over 25 registration combinations and output as data series (Fig. 3). Each data series was inspected for tracking failure using video playback. Any failed tracking data were removed and if all templates failed, the data were not used in the analysis.

Table 1

Participant inclusion and exclusion criteria for repeatability study.

Inclusion criteria	Exclusion criteria
Male and female	Pregnancy
Age 21–51 years	Mental illness
Able to understand written information	Poor understanding of English
Willing to participate	Recent abdominal or pelvic surgery
Able to freely give informed consent	Previous mid-lumbar spinal surgery
Menstruation within last 28 days, or evidence of contraceptive use, or sterility (females)	Body mass index (BMI) > 31
	Medical radiation exposure in the past 2 years with a dose of greater than 8 mSv (defined as CT scan of chest, abdomen or pelvis or interventional procedures under radiological control, i.e. angiography)
Consent to GP being informed of inclusion in study	Current involvement in any other research study
Able to tolerate 80 ° of flexion-extension passive trunk motion	Hyper-mobility syndrome
	Pathology such as fracture, infection, neoplasm
	Spinal stenosis
	Spondylolisthesis
	Radicular pain
	Litigation or compensation pending



Fig. 2. Diagram of patient positioning for fluoroscopic imaging (Ortho Kinematics Inc., with permission).




Fig. 4. Graphical representation of two lumbar vertebrae undergoing extension in the sagittal plane with a four-point reference template marked on the corner of each vertebra to calculate the bisectrix. The bisectrix is to be used as a basis of calculation of translation changes.

2.4. Translation calculation

Frobins method [21] for calculating translation (shown in Figs. 4 and 5 A and B) is based on landmarks identified on the vertebral body 'corners'. Vertebral midlines (Fig. 4) are defined as lines passing through the midpoints between corners 1–2 and 3–4 respectively. The average gradient and y axis crossover of the two mid- lines are calculated for a vertebral pair. The resultant line is called the bisectrix and normally passes through the inter-vertebral disc space. Using the method depicted in Fig. 5, a line is drawn from the centre of each vertebra to the coinciding bisectrix. These lines




Fig. 3. Example of translation data for extension at L5-S1 (live participant). Solid line shows filtered average of 25 tracking. Shaded area represents all data.

intersect the bisectrix at 90° to the bisectors' gradient. Translation was calculated as the distance along the bisectrix between the points at which these two lines independently cross the bisectrix (Fig. 5). To standardise this measurement this is given as a proportion of the mean vertebral body depth of the superior vertebra, where 1 VBU (vertebral body unit) is the mean of the upper and lower vertebral body end plate depth of the superior vertebra. For the in vivo studies VBUs were converted to millimetres based on a standard vertebral depth of 35 mm and for the specimens by their actual measurement.




Fig. 5. ((A) and (B)) Depiction of translation measurement calculation between two adjacent lumbar vertebrae in (A) full extension (B) full flexion.




Fig. 6. ((A) and (B)) Examples of computer-generated measurements of: (A) FCR in fixed centre specimen, (B) translation in movable centre specimen.

2.5. FCR calculation

The FCR position and distance from the posterior superior corner of the inferior vertebral body was calculated by finding the least squares solution between the four corners and the corresponding co-ordinates on the subsequent image [22] (Fig. 6 A and B). The four corner reference template positions for two adjacent vertebrae were taken and re-positioned so that the inferior vertebral position was superimposed. From these coordinate positions, the centre of rotation between the two images was calculated by finding the least squares solution between each of the four corners and their partners from the second image. The least squares solution was taken as described by McCane et al. [22] which gives the Matlab script used to execute this calculation. The positions at which each of these least square solutions meet was taken as the FCR for those two vertebrae between those two images. The axis of rotation was then displayed relative to the inferior vertebra in a pair as a function of the four-corner template on the inferior vertebra. The superior-posterior corner of the inferior vertebra was taken as the origin for this reference field where the X -axis is along the template on the superior vertebral border and the Y - axis perpendicular to the X -axis passing through the origin. The unit of distance used was the proportion of the average vertebral body depth of superior vertebra (due to the non-uniform shape of the sacral template) where the origin of this co-ordinate system is the anterior-superior corner of the inferior vertebra. FCR positional data were calculated at the maximum rotation angle between any two template positions where the inter- vertebral angle was greater than 5 °as a cut-off—as when inter- vertebral rotation interval decreases, the variation in FCR position increases. This is a systematic error due to the way in which the FCR positions are calculated. FCR was measured continuously between the first frame of the image sequence and the image frame where angular rotation was at its maximum $\pm 0.5^\circ$. The limit of $\pm 0.5^\circ$ was selected as this was the increment through which the tracking templates rotated when calculating vertebral body position within each image. The results were taken as the average position of the FCR in X and Y co-ordinates over the 5 tracking's.

Table 2

RMS differences between reference and measured translation and FCR locations.

	Fixed specimen			Translating specimen		
	VBU	mm	95% LoA (VBU)	VBU	mm	95% LoA (VBU)
Translation	0.004	0.10	0.001 to 0.006	0.062	2.16	0.055 to 0.070
FCR x	0.009	0.25	-0.017 to 0.018	–	–	–
FCR y	0.014	0.40	-0.028 to 0.005	–	–	–

2.6. Statistical analysis

For the accuracy study, 10 sets of markings were performed for each specimen. Measured translation was compared with zero translation reference data in the fixed centre specimen (end plate depth 28.77 mm) and with translation across 50% of the inferior end plate (depth 34.66 mm) in the moveable centre specimen. Dis- agreement was expressed as the root-mean-square (RMS) differences between measured and reference values for both translation and FCR. 95% limits of

agreement (LoA) were calculated and expressed in VBU [23]. For the repeatability studies, 4 intervertebral levels (L2-S1) were analysed for both flexion and extension translation for each of the 20 participants. For FCR location, data were removed from FCR analysis when rotation did not reach 5°. This range has been suggested as the lowest over which intervertebral FCRs should be calculated from radiographs without unacceptable error [24]. Therefore, in anticipation that not all levels would reach the necessary 5°, the levels were pooled to give a maximum possible 80 observations for each of flexion and extension. Intra- and inter- observer reliability were expressed as intraclass correlation coefficients (ICC consistency 3,1) using adjectives proposed by Shrout and revised from the original scale of Landis and Koch [25,26]. In the Shrout scale, reliability as denoted by an ICC of 0.00–0.01 is considered as “virtually none”, 0.11–0.40 “slight”, 0.41–0.60 “fair”, 0.61–0.80 “moderate” and 0.81–1.00 “substantial”. 3.

Results

3.1. Accuracy

The proportion of vertebral body depth that was translated in the moveable centre specimen as measured by the actuator motor was 0.52 VBU (17.95 mm). Table 2 shows the RMS differences and 95% LoAs between the reference and measured translation and FCR locations. For the fixed centre of rotation specimen, the average discrepancy (RMS) in translation range between reference and image data was 0.004 VBU (0.10 mm) (LoA 0.01 mm). For the translating specimen, the discrepancy when the superior vertebra was translated across 50% of the end-plate of the lower one was 0.062 VBU (2.16 mm) (LoA 0.52 mm). For FCR, the RMS x and y co-ordinate location differences between the reference and measured locations in the fixed centre specimen were 0.009 VBU (x) or 0.25 mm (LoA 1.30 mm) and for 0.014 VBU (y) or 0.40 mm (LoA 1.20 mm). (Table 2). Bland–Altman plots for these are shown in Fig. 7 (A)–(D).

3.2. Repeatability

The participant sample was made up of 9 females and 11 males aged 26–46 (mean age 35.7, SD 7.20). Their mean body mass index was 24.71 (SD 2.22). Between 6 and 14 observations for each level in the 20 subjects were visible and tracked successfully for translation. Not all levels and directions were visible or trackable in all subjects. Artefacts due to the movement of bowel gas across images and tall patients whose upper vertebral levels did not fit the image field were the main causes of this. Intra- and inter-observer repeatability for each intervertebral level are shown in Table 3. All levels and directions showed at least fair agreement and reliability. The best agreement was between observers at L2-3 in extension (SEM = 0.17 mm) and the worst within observers at L5-S1 in extension (SEM = 1.14 mm). The best reliability was within observers at L2-3 in flexion (ICC = 0.998 (0.958–0.997)) and the worst within observers at L3-4 in flexion (ICC = 0.533 (0.406–0.849)). Repeatability results for FCR are shown in Table 4. Five degrees of rotation was reached by 30 intervertebral pairs. For both translation and FCR location, within observer disagreement did not exceed 2 mm for either flexion or extension. Inter-observer disagreement was high for FCR y in extension (5.67 mm). All directions otherwise showed moderate to substantial reliability, the smallest ICC being 0.621 (0.429–0.813) for FCR x flexion between observers.




Fig. 7. ((A)–(D)) Bland–Altman plots: (A) Translation in fixed centre specimen. (B) Translation in movable centre specimen. (C) FCR x in fixed centre specimen. (D) FCR y in fixed centre specimen.

Table 3

Intra and interobserver repeatability of translation by level and direction.

Level	Flexion						Extension					
	Intraobserver			Intraobserver			Intraobserver			Intraobserver		
	n	SEM (mm)	ICC (95%CI)	n	SEM (mm)	ICC (95%CI)	n	SEM (mm)	ICC (95%CI)	n	SEM (mm)	ICC (95%CI)
L2-3	11	0.18	0.988 (0.958–0.997)	11	0.51	0.865 (0.499–0.964)	7	0.21	0.935 (0.671–0.989)	6	0.17	0.932 (0.514–0.990)
L3-4	14	0.43	0.533 (0.406–0.849)	14	0.46	0.570 (–0.339 to 0.862)	13	0.40	0.742 (0.185–0.920)	12	0.35	0.809 (0.337–0.945)
L4-5	11	0.39	0.853 (0.483–0.947)	11	0.62	0.700 (–0.115 to 0.919)	10	0.56	0.899 (0.619–0.975)	7	0.65	0.916 (0.512–0.982)
L5-S1	13	0.77	0.828 (0.456–0.947)	12	0.75	0.844 (0.458–0.955)	10	1.14	0.644 (–0.344 to 0.910)	8	0.64	0.910 (0.553–0.931)

Table 4

Intra and interobserver repeatability of FCR location (pooled data).

	Flexion						Extension					
	Intraobserver			Intraobserver			Intraobserver			Intraobserver		
	n	SEM (mm)	ICC (95%CI)	n	SEM (mm)	ICC (95%CI)	n	SEM (mm)	ICC (95%CI)	n	SEM (mm)	ICC (95%CI)
IAR x	30	1.72	0.816 (0.678–0.953)	24	2.03	0.621 (0.429–0.813)	21	1.82	0.852 (0.680–1)	21	1.19	0.876 (0.727–1)
IAR y	30	1.75	0.626 (0.421–0.830)	24	1.86	0.690 (0.497–0.882)	21	1.51	0.999 (0.833–1)	21	5.67	0.878 (0.659–1)

4. Discussion

Where mechanical impairment of intervertebral motion in the spine is at issue, its assessment will depend on the availability of technology with which to perform standardised measurements in patients during motion and to provide reference values and error estimates for the various parameters. This study is the first to assess the accuracy and level by level repeatability of the measurement of sagittal plane translation and FCR location from moving vertebral images using low dose 2-D QF. Its results indicate where the current strengths and weaknesses in the technique lie when reporting results of patient studies to clinicians. The accuracy of techniques for radiographic measurement of intervertebral kinematics has been determined using calibration models for roentgen stereo photogrammetry (which although highly invasive, is sometimes considered the gold standard), biplanar radiography and QF [10,15,27,28]. In this study, idealised conditions were also avoided by degrading the images with animal soft tissue and in the upright position, although it is not uncommon for such studies to be undertaken with no loading or in an animal model with no tissue degradation [16,29]. In this study, we compensated for radiographic image distortion using distortion-compensated roentgen analysis and used an image intensifier that incorporated automatic distortion correction [21]. Measurement is virtually independent of distortion of the radiographic image resulting from central projection, axial rotation, lateral tilt, and off-centre position with an error for translation of between 0.4 and 0.8 mm. Measurement of translation was determined from the vertebral body centres, making it independent of rotation. Previous QF studies have also shown that degrading the alignment by axially rotating it 10 ° out of plane and inclining the X-ray beam inclined 10 ° inferiorly results in minimal loss of accuracy in rotational studies [10]. Thus the technique should be sufficiently accurate to give useful information about ranges and motion patterns. However, this technique is not thought to be possible in scoliotic spines due to failure of image tracking. This study found the current QF method to have fair to substantial repeatability for all levels and directions using the current protocol. It also found acceptable accuracy in vitro for the measurement of FCR location and translation during continuous spinal motion. Reliability was mainly good, but at some levels and directions suggests that training and quality assurance are needed when applying the measurement to comparisons between individuals and reference standards [30].

The inter-observer y -error in determination of FCR in extension (5.67 mm) and the intra-observer ICC (0.644) for extension translation at L5-S1 point to a need for caution. Closer inspection of the data revealed that the former was also greatest at L5-S1, where image quality and consequently co-ordinate placement may be rendered problematical by the super-imposition of the ilia and/or lack of perfect orthogonal alignment of the central X-ray beam with the vertical axis of the vertebrae. Previous work found radiographic positioning to be more important than tracking accuracy as a contributor to the variability in measurement of angular position, but that this does not preclude high repeatability and accuracy of measurement of rotation [19,47]. However, for translation and FCR this may be more critical. FCR was once thought to be promising as a way of assessing abnormal loading during intervertebral motion in patients [31,32] but fell out of favour owing to high errors in measurement and the intrinsic computational errors that occur when rotational range is low [24,33–35]. The suggestion that it might be used to measure stability has therefore also not generally been taken up [14]. However, the present study has shown that despite the use of continuous motion data, as is necessary in patient studies, greater accuracy was achieved for determining the FCR (average error 0.3 mm x , 0.4 mm y) than was found in a previous study with such a specimen that used stepped rotation positions (average error 2 mm) [15]. The repeatability

study utilised information from participants undergoing passive recumbent and not weight bearing motion. It may be thought that weight bearing information would have been preferable to study the repeatability of translation and FCR measurement. However, this would have meant irradiating additional participants to obtain the same data and differences in motion patterns associated with weight bearing should not affect their measurement. Indeed, Wood et al. concluded that the lateral decubitus position was superior for the detection of instability in patients with spondylolisthesis and Yeager et al. used these interchangeably for their repeatability analysis of rotation and translation at pooled levels [36,5]. FCR, at least in the sagittal plane, could therefore be used to inform both patient care and patient-specific mathematical models. However, further studies are needed to establish normative in vivo reference standards at individual levels using QF. It would also be beneficial to explore the effects of spinal geometry and muscle contraction on FCR location, to add coronal plane validation and to confirm whether the FCR locus might be used to assess relationships between structural change and the in vivo biomechanical performance characteristics of discs under load. Finally, rotational cut-offs for accurately locating the FCR should be revisited in the light of the greater standardisation offered by QF protocols. Diagnostic advances in spine biomechanics have also been made using kinetic MRI [37–40] and SPECT-CT imaging [41,42]. However, although kinetic MRI locates points of encroachment on neural tissues and SPECT-CT contributes to the identification of potential sites of pain generation, neither can extract end-range or continuous inter-vertebral motion. In addition, the radiation dosage from SPECT-CT is considerably larger than that of QF. Improvements in repeatability and accuracy are ongoing requirements for any diagnostic test, which means that reference standards will always be imperfect. Validation of QF will therefore require that scientists and practitioners also examine the extent to which test results are meaningful in practice [43]. This may be appreciated from patient register data. In parallel with this, technology development should address any measurement deficiencies.

5. Limitations

Participants with a BMI over 31 or aged over 51 were excluded from the study and none had osteoporosis, osteoarthritic change, vertebral deformities or curvatures; which may precipitate tracking failures. In the accuracy study, the translation error was considerably higher (2.10 mm) in the translating specimen than in the fixed specimen (0.10 mm). This may have been due to the resolution of the actuator motor in the latter (0.01 mm), or by a small amount of out of plane motion due to imperfections in the mechanical linkage of this specimen. However, this discrepancy is well below the generally accepted cut off of 4 mm for excessive translation [44–47]. Distortion that changes during motion is not correctable if the templates that track the images from frame to frame do not change to accommodate it. In the future, this could be provided by adaptations to the tracking codes [8]. The US versions of this technology image the upper and lower lumbar levels separately to minimise out of plane images and ensure inclusion of all lumbar levels. While this increases the X-ray dose, it also makes for better reliability in the measurement of translation than was found here [5]. Future studies of accuracy and repeatability are needed to substantiate the present work. These could use a larger number of examiners, a range of rotational angles for FCR accuracy and a more elaborate calibration set up that combines rotation and translation. A larger number of human participants would overcome the problem of low angles of rotation and enable determination of the level by level repeatability of FCR location at 5° and above. For example, poorer agreement was found at L5-S1 than other levels, possibly owing to lower image quality resulting from superimposition of both ilia on the vertebral images. 6.

Conclusion

Quantitative fluoroscopy was found to have a high level of accuracy as well as moderate to substantial observer agreement and reliability for the measurement of FCR and translation. Exceptions were in the reliability of measuring translation at L3-4 and agreement between observers in locating the FCR in extension. The development of reference standards and analysis quality assurance measures will be essential for optimal clinical use [6].

Conflict of interest

The authors have performed research for the Ortho Kinematics Company, which is commercialising a version of this technology in the United States.

Sources of funding

No external funding was obtained for this research.

Ethical approval

Ethical approval was given by the National Research Ethics Service (REC reference 0/H0502/99).

Acknowledgements

This article presents independent research and the views expressed are those of the author(s) alone. The authors wish to acknowledge the contributions of the participants and of Fiona Mellor for sharing some of her participant image sequences and contributing to the participant image registration.

References

- [1] Deyo RA , McNeish LM , Cone RO 3rd . Observer variability in the interpretation of lumbar spine radiographs. *Arthritis Rheum* 1985;1528:1066–70 .
- [2] Shaffer WO , Spratt KF , Weinstein JD , Lehmann TR , Goel V . Volvo award in clinical sciences: the consistency and accuracy of roentgenograms for measuring sagittal translation in the lumbar vertebral motion segment: an experimental model. *Spine* 1990;15:741–50 .
- [3] Penning L, Wilmink JT, Van Woerden HH. Inability to prove instability: a critical appraisal of clinical-radiological flexion-extension studies in lumbar disc degeneration. *Diagn Imaging Clin Med* 1984;53:186–92 .
- [4] Cakir B , Richter M , Kafer W , Wieser M , Puhl W , Schmidt R . Evaluation of lumbar spine motion with dynamic x-ray: a reliability analysis. *Spine* 2006;31:1258–64 .
- [5] Yeager MS , Cook DJ , Cheng BC . Reliability of computer-assisted lumbar intervertebral measurement using a novel vertebral motion analysis system. *Spine J* 2014;14:274–81 .
- [6] Leone A , Guglielmi G , Cassar-Pullicino VN , Bonomo L . Lumbar intervertebral instability: a review. *Radiology* 2007;245:62–77

- [7] Breen AC , Allen R , Morris A . An image processing method for spine kinematics-preliminary studies. *Clin Biomech* 1988;3:5–10 .
- [8] Breen AC , Teyhan DS , Mellor FE , Breen AC , Wong KWN , Deitz A . Measurement of intervertebral motion using quantitative fluoroscopy: report of an international forum and proposal for use in the assessment of degenerative disc disease in the lumbar spine. *Adv Orthop* 2012;2012:1–10 .
- [9] Mellor FE , Thomas P , Thompson P , Breen AC . Proportional lumbar spine inter-vertebral motion patterns: a comparison of patients with chronic non-specific low back pain and healthy controls. *Eur Spine J* 2014;23:2059–67 .
- [10] Breen A , Muggleton J , Mellor F . An objective spinal motion imaging assessment (OSMIA): reliability, accuracy and exposure data. *BMC Musculoskelet Disord* 2006;7:1–10 .
- [11] Breen A , Mellor F , Breen A . Lumbar intervertebral motion in vivo: a preliminary comparison of recumbent and weight bearing motion patterns in adult males. *Bone Joint J* 2013;95-B:20 .
- [12] Branney J , Breen AC . Does inter-vertebral range of motion increase after spinal manipulation? A prospective cohort study. *Chiropractic Manual Ther* 2014;22:24 .
- [13] Iguchi T , Kanemura A , Kasahara K , Sato K , Kurihara A , Yoshiya S , et al. Lumbar instability and clinical symptoms. Which is the more critical factor for symptoms: sagittal translation or segment angulation. *J Spinal Disord Tech* 2004;17:284–90 .
- [14] Bogduk N , Amevo B , Pearcy M . A biological basis for instantaneous centres of rotation of the vertebral column. *Proc Instn Mech Eng* 1995;209:177–83 .
- [15] Cerciello T , Romano M , Bifulco P , Cesarelli M , Allen R . Advanced template matching method for estimation of intervertebral kinematics of lumbar spine. *Med Eng Phys* 2011;33:1293–302 .
- [16] Wang S , Passias P , Li G , Li G , Wood K . Measurement of vertebral kinematics using non-invasive image matching method-validation and application. *Spine* 2008;33:E355–61 .
- [17] Lin H , Wang S , Tsai T-Y , Li G , Kwon Y-M . In-vitro validation of a non-invasive dual fluoroscopic imaging technique for measurement of the hip kinematics; 2013;35:411–16 .
- [18] Mellor FE , Thomas P , Breen AC . Quantitative fluoroscopy for investigating in vivo kinematics of the lumbar spine; radiation dose compared to lumbar spine radiographs with suggestions for further dose reduction. *Radiography* 2014;20:251–7 .
- [19] Dvorak J , Panjabi MM , Chang DG , Theiler R , Grob D . Functional radiographic diagnosis of the lumbar spine: flexion-extension and lateral bending. *Spine* 1991;16:562–71 .
- [20] Muggleton JM , Allen R . Automatic location of vertebrae in digitised video fluoroscopic images of the lumbar spine. *Med Eng Phys* 1997;19:77–89 .
- [21] Frobin F , Brinckmann P , Lievseth G , Biggemann M , Reikeras O . Precision measurement of segmental motion from flexion-extension radiographs of the lumbar spine. *Clin Biomech* 1996;11:457–65 .

- [22] McCane B , Abbott JH , King T . On calculating the finite centre of rotation for rigid planar motion. *Med Eng Phys* 2005;27:75–9 .
- [23] Bland JM , Altman DG . Statistical methods for assessing agreement between two methods of clinical measurement. *The Lancet* 1986:307–10 .
- [24] Pearcy M , Bogduk N . Instantaneous axes of rotation of the lumbar intervertebral joint. *Spine* 1988;13:1033–41 .
- [25] Shrout PE . Measurement reliability and agreement in psychiatry. *Stat Methods Med Res* 1998;7:301–17 .
- [26] Landis JR , Koch GG . The measurement of observer agreement for categorical data. *Biometrics* 1977;33:159–74 .
- [27] Pearcy MJ . Stereoradiography of lumbar spine motion. *Acta Orthop Scand* 1985;56:1–45 .
- [28] Selvik G . Roentgen stereo photogrammetry: a method for the study of the kinematics of the skeletal system. *Acta Orthop Scand* 1989;232(Suppl):1–51 .
- [29] Zuhlke T , Fine J , Haughton VM , Anderson PA . Accuracy of dynamic computed tomography to calculate rotation occurring at lumbar spinal motion segments. *Spine* 2009;34:E215–18 .
- [30] Knottner J , Audige L , Brorson S , Donner A , Gajewski BJ , Hrobjartsson A , et al. Guidelines for reporting reliability and agreement studies (GRRAS) were proposed. *J Clin Epidemiol* 2011;64:96–106 .
- [31] Gertzbein SD , Seligman J , Holtby K . Centrode patterns and segmental instability in degenerative disc disease. *Spine* 1985;10:257–61 .
- [32] Seligman JV , Gertzbein SD , Tile M , Kapasouri A . Computer analysis of spinal segment motion in degenerative disc disease with and without axial loading. *Spine* 1984;9:566–73 .
- [33] Panjabi MM . Centres and angles of rotation of body joints: a study of errors and optimization. *J Biomech* 1979;12:911–20 .
- [34] Soudan K , Van Audekercke R , Martens M . Methods, difficulties and inaccuracies in the study of human joint kinematics and pathokinematics by the instant axis concept, example: the knee joint. *J Biomech* 1979;12:27–33 .
- [35] Panjabi MM , Goel VK , Takata K . Physiologic strains in the lumbar spinal ligaments. *Spine* 1982;7:192–203 .
- [36] Wood KB , Popp CA , Transfeldt EE , Geissele AE . Radiographic evaluation of in- stability in spondylolisthesis. *Spine* 1994;19:1697–703 .
- [37] Hirasawa Y , Bashir WA , Smith FW , Magnusson ML , Pope MH , Takahashi K . Postural changes of the dural sac in the lumbar spines of asymptomatic individuals using positional stand-up magnetic resonance imaging. *Spine* 2007;32:E136–40 .

- [38] Kong MH , Hymanson HJ , Song KY , Chin DK , Cho YE , Yoon DH , et al. Kinetic magnetic resonance imaging analysis of abnormal segmental motion of the functional spine unit. *J Neurosurg Spine* 2009;10:357–65 .
- [39] Tan Y , Aghdasi BG , Montgomery SR , Inoue H , Lu C , Wang JC . Kinetic magnetic resonance imaging analysis of lumbar segmental mobility in patients without significant spondylosis. *Eur Spine J* 2012;21:2673–9 .
- [40] Lao L , Daubs MD , Scott TP , Lord EL , Cohen JR , Tin R , et al. Effect of disc degeneration on lumbar segmental mobility analyzed by kinetic magnetic resonance imaging. *Spine* 2015;40:316–22 .
- [41] Harisankar CNB , Mittal BR , Bhattacharya A , Singh P , Sen R . Utility of single photon emission computed tomography/computed tomography imaging in evaluation of chronic low back pain. *Indian J Null Med* 2012; 27:156–63.
- [42] Matar HE , Navalkissoor S , Berovic M , Shetty R , Garlick N , Casey ATH , et al. Is hybrid imaging (SPECT/CT) a useful adjunct in the management of suspected facet joints arthropathy? *Int Orthop* 2013;37:865–70 .
- [43] Breen AC , Morris A . Lumbar spine instantaneous centres of rotation determined by digital video-fluoroscopy. London: Society for Back Pain Research Annual Scientific Meeting. St Mary's Hospital; 1989 .
- [44] Dupuis PR , Yong-Hing K , Cassidy JD , Kirkaldy-Willis WH . Radiologic diagnosis of degenerative lumbar spinal instability. *Spine* 1985;10:262–76 .
- [45] Boden SD , Wiesel SW . Lumbosacral segmental motion in normal individuals—have we been measuring instability properly? *Spine* 1990;15:571–6 .
- [46] Morgan FP , King I . Primary instability of lumbar vertebrae as a common cause of low back pain. *J Bone Joint Surg* 1957; 39B:6–22. [47] Hanley EN . The indications for lumbar spinal fusion with and without instrumentation. *Spine* 1995;20:143–53 . Please